



STEM CELL LABORATORY (STCL)



DOCUMENT NUMBER: STCL-FORM-022

DOCUMENT TITLE:

Centrifuge Bi-Annual Verification of Function

DOCUMENT NOTES:

8B.612

Document Information

Revision: 03

Vault: STCL-Form-rel

Status: Release

Document Type: STCL FORM

Date Information

Creation Date: 14 Jun 2013

Release Date: 01 Jul 2013

Effective Date: 01 Jul 2013

Expiration Date:

Control Information

Author: WATE02

Owner: WATE02

Previous Number: STCL-FORM-022 Rev 02

Change Number: QSU-CCR-105

Stem Cell Laboratory

Equipment: _____
 SN: _____
 Location: _____
 Year: _____

CENTRIFUGE
BI-ANNUAL VERIFICATION OF FUNCTION
 (Verify all RPM setting and times used routinely)

* = Record problem/corrective action on reverse side

	JAN-JUNE		JULY-DEC	
RPM CHECK	Speed		Speed	
	<u>Indicator</u>	<u>Tach</u>	<u>Indicator</u>	<u>Tach</u>
TOLERANCE LIMITS				
± 100 RPMs	_____	_____	_____	_____
	_____	_____	_____	_____
	_____	_____	_____	_____
	_____	_____	_____	_____
Tech Initials/Date				
TIMER CHECK	<u>Time</u>	<u>Stopwatch</u>	<u>Time</u>	<u>Stopwatch</u>
TOLERANCE LIMITS ± 30 seconds	_____	_____	_____	_____
	_____	_____	_____	_____
	_____	_____	_____	_____
	_____	_____	_____	_____
Initials/Date				
Reviewed By/ Date				

Stem Cell Laboratory

Equipment: Sorval BP
 SN: SBP2556248
 Location: Processing
 Year: 2010

CENTRIFUGE—EXAMPLE
BI-ANNUAL VERIFICATION OF FUNCTION
 (Verify all RPM setting and times used routinely)

* = Record problem/corrective action on reverse side

	JAN-JUNE		JULY-DEC	
RPM CHECK	Speed		Speed	
	<u>Indicator</u>	<u>Tach</u>	<u>Indicator</u>	<u>Tach</u>
TOLERANCE LIMITS	<u>700</u>	<u>710</u>	<u> </u>	<u> </u>
± 100 RPMs	<u>1200</u>	<u>1190</u>	<u> </u>	<u> </u>
	<u>1800</u>	<u>1810</u>	<u> </u>	<u> </u>
	<u>2500</u>	<u>2550</u>	<u> </u>	<u> </u>
Tech Initials/Date	HH 1/25/2010			
TIMER CHECK	<u>Time</u>	<u>Stopwatch</u>	<u>Time</u>	<u>Stopwatch</u>
	<u>10</u>	<u>10min 15sec</u>	<u> </u>	<u> </u>
TOLERANCE LIMITS ± 30 seconds	<u>15</u>	<u>15min 3sec</u>	<u> </u>	<u> </u>
	<u>20</u>	<u>19min 55sec</u>	<u> </u>	<u> </u>
	<u> </u>	<u> </u>	<u> </u>	<u> </u>
Initials/Date	HH 1/25/2010			
Reviewed By/ Date				

Signature Manifest**Document Number:** STCL-FORM-022**Revision:** 03**Title:** Centrifuge Bi-Annual Verification of Function**STCL-FORM-022 Centrifuge Bi-Annual****Author Approval**

Name/Signature	Title	Date	Meaning/Reason
Barbara Waters-Pick (WATE02)		19 Jun 2013, 01:12:13 PM	Approved

Medical Director Approval

Name/Signature	Title	Date	Meaning/Reason
Joanne Kurtzberg (KURTZ001)		19 Jun 2013, 03:05:11 PM	Approved

QA Approval

Name/Signature	Title	Date	Meaning/Reason
Linda Sledge (SLEDG006)		20 Jun 2013, 08:26:22 AM	Approved

Document Release

Name/Signature	Title	Date	Meaning/Reason
Sandy Mulligan (MULLI026)		01 Jul 2013, 02:25:07 PM	Approved

Notification

Name/Signature	Title	Date	Meaning/Reason
Linda Sledge (SLEDG006)		01 Jul 2013, 02:25:08 PM	Email Sent
Sharon Hartis (SH259)		01 Jul 2013, 02:25:08 PM	Email Sent
Betsy Jordan (BJ42)		01 Jul 2013, 02:25:08 PM	Email Sent
Barbara Waters-Pick (WATE02)		01 Jul 2013, 02:25:08 PM	Email Sent